

# BIOMETER OAS 2000



## Specifications

<b>Probe</b>	10MHz
<b>Storage</b>	10groups
<b>Accuracy</b>	0.05mm
<b>IOL Calculation</b>	SRK-II,SRK/T Hoffer-Q, Holladay, SCDK , etc
<b>Measure</b>	ACD, Lens, Vitreous, Axial Length, and its Average
<b>Modes</b>	Manual, Automatic
<b>Printer</b>	Inner therm Printer
<b>Power</b>	AC 220V±10% 50Hz
<b>Input</b>	Touch screen
<b>Display Monitor</b>	High speed LCD
<b>Report Data</b>	Patient's name, ID, Right/Left, Doctor's name, Date, IOL Table